

Certificate of Analysis and Release

FertiCult Flushing Medium

| | | |
|------------------|------------|------------|
| Batchnumber: | | FP17FL17 |
| Production date: | DD/MM/YYYY | 20/04/2017 |
| Expiry date: | DD/MM/YYYY | 31/10/2018 |
| Total volume: | | 124000 ml |

| | <i>Results</i> | <i>Specifications</i> |
|-----------------------------|----------------|-----------------------|
| pH: | 7,39 | 7.30-7.60 |
| Osmolality: | 282 | 270-290mOsm/kg |
| Bacterial screen: | PASSED | PASSED |
| Fungal screen: | PASSED | PASSED |
| Endotoxin: | < 0,05 EU/ml | < 0,25EU/ml |
| Concentration HSA: | 3,976 | 3.2 - 4.8 g/liter |
| Mouse embryo test (1-cell): | 95 % | ≥ 80% |

Product released by Justine Van Renterghem on 05/05/2017 (DD/MM/YYYY)

