

Certificate of Analysis and Release

FertiCult Flushing Medium

Batchnumber:		FP17FL22
Production date:	DD/MM/YYYY	06/06/2017
Expiry date:	DD/MM/YYYY	31/12/2018
Total volume:		130000 ml

	<i>Results</i>	<i>Specifications</i>
pH:	7,47	7.30-7.60
Osmolality:	280	270-290mOsm/kg
Bacterial screen:	PASSED	PASSED
Fungal screen:	PASSED	PASSED
Endotoxin:	< 0,05 EU/ml	< 0,25EU/ml
Concentration HSA:	3,80	3.2 - 4.8 g/liter
Mouse embryo test (1-cell):	95%	≥ 80%

Product released by Justine Van Renterghem on 21/06/2017 (DD/MM/YYYY)

