

## Certificate of Analysis and Release

### FertiCult Flushing Medium

Batchnumber:		FP17FL29PRG
Production date:	DD/MM/YYYY	25/07/2017
Expiry date:	DD/MM/YYYY	31/01/2019
Total volume:		21825 ml

	<i>Results</i>	<i>Specifications</i>
pH:	7,44	7.30-7.60
Osmolality:	275	270-290mOsm/kg
Bacterial screen:	PASSED	PASSED
Fungal screen:	PASSED	PASSED
Endotoxin:	< 0,05 EU/mL	< 0,25EU/ml
Concentration HSA:	3,444	3.2 - 4.8 g/liter
Concentration gentamicin:	9,52	5 - 15 mg/liter
Mouse embryo test (1-cell):	100 %	≥ 80%

Product released by Eileen Vanderplancke on 09/08/2017 (DD/MM/YYYY)

*Eileen Vanderplancke*