

## Certificate of Analysis and Release

### FertiCult Flushing Medium

Batchnumber:		FP17FL34
Production date:	DD/MM/YYYY	28/08/2017
Expiry date:	DD/MM/YYYY	28/02/2019
Total volume:		13750 ml

	<i>Results</i>	<i>Specifications</i>
pH:	7,42	7.30-7.60
Osmolality:	281	270-290mOsm/kg
Bacterial screen:	PASSED	PASSED
Fungal screen:	PASSED	PASSED
Endotoxin:	< 0,05 EU/mL	< 0,25EU/ml
Concentration HSA:	3,744	3.2 - 4.8 g/liter
Mouse embryo test (1-cell):	100 %	≥ 80%

Product released by Eileen Vanderplancke on 12/09/2017 (DD/MM/YYYY)

Eileen Vanderplancke

*Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HCV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.*