

Certificate of Analysis and Release Product: FertiVit

Batchnumber :		FP17FVW05
Production date:	DD/MM/YYYY	15/06/2017
Expiry date:	DD/MM/YYYY	30/06/2018
Total batchsize:		150 Kits
FertiVit W1	Result	Specification
pH:	7,21	7.2 - 7.4
Sterility test:	PASSED	PASSED
Endotoxin:	< 0,1 EU/ml	< 0,25EU/mL
Osmolality:	1671	-
Concentration HSA:	16,5	13,9 - 20,9 g/liter
FertiVit W2		
pH:	7,26	7.2 - 7.4
Sterility test:	PASSED	PASSED
Endotoxin :	< 0,1 EU/ml	< 0,25EU/mL
Osmolality:	1244	- 444 040 = //:
Concentration HSA:	17,5	14,4 - 21,6 g/liter
FertiVit W3	7.07	70.74
pH:	7,27	7.2 - 7.4
Sterility test:	PASSED	PASSED
Endotoxin: Osmolality :	< 0,1 EU/mI 822	< 0,25EU/mL 805-850mOsm/kg
Concentration HSA:	17,8	14,9 - 22,3 g/liter
	17,0	14,9 - 22,3 g/iitei
FertiVit W4	7.00	70 74
pH:	7,32	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin:	< 0,1 EU/ml 547	< 0,25EU/mL 535-565mOsm/kg
Osmolality:		
Concentration HSA:	19,9	15,4 - 23,1 g/liter
FertiVit W5	7.00	70 74
pH:	7,30	7.2 - 7.4
Sterility test : Endotoxin:	PASSED < 0,1 EU/ml	PASSED < 0,25EU/mL
Osmolality:	< 0,1 E0/IIII 414	< 0,25E0/IIIL 405-435mOsm/kg
Concentration HSA:		•
	20,8	15,7 - 23,6 g/liter
FertiVit W6		70 74
pH:	7,31	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin: Osmolality:	< 0,1 EU/ml 282	< 0,25EU/mL
Concentration HSA:	20,1	270-290mOsm/kg
		16,0 - 24,0 g/liter
Mouse Embryo Assay (blastocyst after 96h):	85,71	≥ 80%

Product released by Lenie Bocklandt on 25/07/2017 (DD/MM/YYYY)

Caution: All blood products should be treated as potentially infectious. Source material from which this product was derived was found negative when tested for antibodies to HIV, HBc, HCV, and non-reactive for HbsAg, HCV RNS and HIV-1 RNA and syphillis. No known test methods can offer assurance that products derived from human blood will not transmit infectious agents.

