

Certificate of Analysis and Release

Product: SpermMar IgG

| | | |
|--------------------------|----------|------------|
| Batchnumber : | | FP17G04 |
| Production date : | DD/MM/YY | 6/04/2017 |
| Expiry date : | DD/MM/YY | 31/10/2018 |
| Number of kits in batch: | | 532 kits |

Chemical analysis

| | | |
|----------------|--|--------|
| pH: | | 8,19 |
| Osmolality: | | 357 |
| Concentration: | | PASSED |

Functional analysis

| | | |
|-------------------|--|--------|
| Positive control: | | PASSED |
| Negative control: | | PASSED |

Product released by Justine Van Renterghem on 14/04/2017 (DD/MM/YYYY)

A handwritten signature in blue ink, appearing to be "Justine Van Renterghem", written over a horizontal line.