

## Certificate of Analysis and Release

Product: Hypo-osmotic swelling test

Batchnumber :		FP17HOS02
Production date :	DD/MM/YYYY	20/04/2017
Expiry date :	DD/MM/YYYY	30/04/2018
Batch size:		61 kits

	<i>Result</i>	<i>Specification</i>
Osmolality:	166 mOsm/kg	160-170mOsm/kg
pH:	7,30	7.2-7.6
Functionality test:	PASSED	PASSED

Product released by Justine Van Renterghem on 05/05/2017 (DD/MM/YYYY)

A handwritten signature in blue ink, appearing to be "Justine Van Renterghem", written over a horizontal line.