

Certificate of Analysis and Release

Product: Hypo-osmotic swelling test

Batchnumber :		FP17HOS03
Production date :	DD/MM/YYYY	18/08/2017
Expiry date :	DD/MM/YYYY	31/08/2018
Batch size:		60 kits

	<i>Result</i>	<i>Specification</i>
Osmolality:	165 mOsm/kg	160-170mOsm/kg
pH:	7,31	7.2-7.6
Functionality test:	PASSED	PASSED

Product released by Eileen Vanderplancke on 04/09/2017 (DD/MM/YYYY)

Eileen Vanderplancke