

Certificate of Analysis and Release

Hyaluronidase in FertiCult Flushing medium

Batchnumber:		FP17HY05
Production date:	DD/MM/YYYY	17/03/2017
Expiry date:	DD/MM/YYYY	31/03/2018
Total volume:		3300 ml

	<i>Results</i>	<i>Specifications</i>
pH:	7,44	7.30-7.60
Osmolality:	282	270-290mOsm/kg
Sterility (fungal and bacterial sterility test):	PASSED	PASSED
Endotoxin:	0,117 EU/ml	< 1 EU/ml
Concentration HSA:	3,928	3.2 - 4.8 g/liter
Enzyme activity:	102,59	40 - 160 IU/ml
Mouse embryo test (1-cell):	100%	≥ 80%

Product released by Justine Van Renterghem on 03/04/2017 (DD/MM/YYYY)



Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HCV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.