

Certificate of Analysis and Release

Hyaluronidase in FertiCult Flushing medium

Batchnumber:		FP17HY07
Production date:	DD/MM/YYYY	19/04/2017
Expiry date:	DD/MM/YYYY	30/04/2018
Total volume:		3000 ml

	<i>Results</i>	<i>Specifications</i>
pH:	7,44	7.30-7.60
Osmolality:	280	270-290mOsm/kg
Sterility (fungal and bacterial sterility test):	PASSED	PASSED
Endotoxin:	0,165	< 1 EU/ml
Concentration HSA:	3,835	3.2 - 4.8 g/liter
Enzyme activity:	40,65	40 - 160 IU/ml
Mouse embryo test (1-cell):	100 %	≥ 80%

Product released by Justine Van Renterghem on 04/05/2017 (DD/MM/YYYY)

