

## Certificate of Analysis and Release

### Hyaluronidase in FertiCult Flushing medium

Batchnumber:		FP17HY08
Production date:	DD/MM/YYYY	11/05/2017
Expiry date:	DD/MM/YYYY	31/05/2018
Total volume:		3800 ml

	<i>Results</i>	<i>Specifications</i>
pH:	7,43	7.30-7.60
Osmolality:	282	270-290mOsm/kg
Sterility (fungal and bacterial sterility test):	PASSED	PASSED
Endotoxin:	0,159	< 1 EU/ml
Concentration HSA:	3,8	3.2 - 4.8 g/liter
Enzyme activity:	86,94	40 - 160 IU/ml
Mouse embryo test (1-cell):	100 %	≥ 80%

Product released by Justine Van Renterghem on 06/06/2017 (DD/MM/YYYY)

