

Certificate of Analysis and Release

Hyaluronidase in FertiCult Flushing medium

Batchnumber:		FP17HY09
Production date:	DD/MM/YYYY	31/05/2017
Expiry date:	DD/MM/YYYY	31/05/2018
Total volume:		4200 ml

	<i>Results</i>	<i>Specifications</i>
pH:	7,46	7.30-7.60
Osmolality:	279	270-290mOsm/kg
Sterility (fungal and bacterial sterility test):	PASSED	PASSED
Endotoxin:	0,178 EU/ml	< 1 EU/ml
Concentration HSA:	3,70	3.2 - 4.8 g/liter
Enzyme activity:	81,52	40 - 160 IU/ml
Mouse embryo test (1-cell):	95,24%	≥ 80%

Product released by Justine Van Renterghem on 15/06/2017 (DD/MM/YYYY)

