

## Certificate of Analysis and Release

### Hyaluronidase in FertiCult Flushing medium

Batchnumber:		FP17HY13
Production date:	DD/MM/YYYY	04/08/2017
Expiry date:	DD/MM/YYYY	31/08/2018
Total volume:		1700 ml

	<i>Results</i>	<i>Specifications</i>
pH:	7,46	7.30-7.60
Osmolality:	279	270-290mOsm/kg
Sterility (fungal and bacterial sterility test):	PASSED	PASSED
Endotoxin:	0,154 EU/mL	< 1 EU/ml
Concentration HSA:	3,942	3.2 - 4.8 g/liter
Enzyme activity:	55,30	40 - 160 IU/ml
Mouse embryo test (1-cell):	100 %	≥ 80%

Product released by Eileen Vanderplancke on 23/08/2017 (DD/MM/YYYY)

*Eileen Vanderplancke*