

## Certificate of Analysis and Release

### Hyaluronidase in FertiCult Flushing medium

Batchnumber:		FP17HY15
Production date:	DD/MM/YYYY	15/09/2017
Expiry date:	DD/MM/YYYY	30/09/2018
Total volume:		1855 ml

	<i>Results</i>	<i>Specifications</i>
pH:	7,49	7.30-7.60
Osmolality:	278	270-290mOsm/kg
Sterility (fungal and bacterial sterility test):	PASSED	PASSED
Endotoxin:	0,203 EU/mL	< 1 EU/ml
Concentration HSA:	3,757	3.2 - 4.8 g/liter
Enzyme activity:	93,73	40 - 160 IU/ml
Mouse embryo test (1-cell):	95,24%	≥ 80%

Product released by Eileen Vanderplancke on 04/10/2017 (DD/MM/YYYY)

*Eileen Vanderplancke*

*Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HCV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.*