

Certificate of Analysis and Release

Hyaluronidase in FertiCult Flushing medium

Batchnumber:		FP17HY18
Production date:	DD/MM/YYYY	30/10/2017
Expiry date:	DD/MM/YYYY	31/10/2018
Total volume:		2685 ml

	<i>Results</i>	<i>Specifications</i>
pH:	7,43	7.30-7.60
Osmolality:	280	270-290mOsm/kg
Sterility (fungal and bacterial sterility test):	PASSED	PASSED
Endotoxin:	0,165 EU/mL	< 1 EU/ml
Concentration HSA:	4,186	3.2 - 4.8 g/liter
Enzyme activity:	114	40 - 160 IU/ml
Mouse embryo test (1-cell):	100 %	≥ 80%

Product released by Eileen Vanderplancke on 14/11/2017 (DD/MM/YYYY)

Eileen Vanderplancke

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HCV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.