

Certificate of Analysis and Release

10% PVP in FertiCult Flushing medium

Batchnumber:		FP17PVP07
Production date:	DD/MM/YYYY	10/05/2017
Expiry date:	DD/MM/YYYY	28/02/2018
Total volume:		640 ml

	<i>Results</i>	<i>Specifications</i>
pH:	7,37	7.20-7.60
Osmolality:	311	280-320 mOsm/kg
Sterility (fungal and bacterial sterility test) :	PASSED	PASSED
Endotoxin:	0,109 EU/mL	< 1 EU/ml
Concentration HSA:	3,70	3.2 - 4.8 g/liter
Viscosity (at 25°C)	260,3	> 170 cP
Mouse embryo test (1-cell):	95 %	≥ 80%

Product released by Lenie Bocklandt on 01/06/2017 (DD/MM/YYYY)



Bocklandt Lenie