

Certificate of Analysis and Release

10% PVP in FertiCult Flushing medium

Batchnumber:		FP17PVP08
Production date:	DD/MM/YYYY	18/05/2017
Expiry date:	DD/MM/YYYY	28/02/2018
Total volume:		720 ml

	<i>Results</i>	<i>Specifications</i>
pH:	7,29	7.20-7.60
Osmolality:	304	280-320 mOsm/kg
Sterility (fungal and bacterial sterility test) :	PASSED	PASSED
Endotoxin:	< 0,05 EU/ml	< 1 EU/ml
Concentration HSA:	3,70	3.2 - 4.8 g/liter
Viscosity (at 25°C)	195,6	> 170 cP
Mouse embryo test (1-cell):	95 %	≥ 80%

Product released by Justine Van Renterghem on 07/06/2017 (DD/MM/YYYY)

