

Certificate of Analysis and Release

10% PVP in FertiCult Flushing medium

Batchnumber:		FP17PVP11
Production date:	DD/MM/YYYY	31/07/2017
Expiry date:	DD/MM/YYYY	30/04/2018
Total volume:		81 ml

	<i>Results</i>	<i>Specifications</i>
pH:	7,32	7.20-7.60
Osmolality:	302	280-320 mOsm/kg
Sterility (fungal and bacterial sterility test) :	PASSED	PASSED
Endotoxin:	< 0,05 EU/mL	< 1 EU/ml
Concentration HSA:	3,648	3.2 - 4.8 g/liter
Viscosity (at 25°C)	238,7	> 170 cP
Mouse embryo test (1-cell):	100 %	≥ 80%

Product released by Eileen Vanderplancke on 16/08/2017 (DD/MM/YYYY)

Eileen Vanderplancke