

Certificate of Analysis and Release

10% PVP in FertiCult Flushing medium

Batchnumber:		FP17PVP13
Production date:	DD/MM/YYYY	29/08/2017
Expiry date:	DD/MM/YYYY	31/05/2018
Total volume:		149 ml

	<i>Results</i>	<i>Specifications</i>
pH:	7,50	7.20-7.60
Osmolality:	301	280-320 mOsm/kg
Sterility (fungal and bacterial sterility test) :	PASSED	PASSED
Endotoxin:	0,101 EU/mL	< 1 EU/ml
Concentration HSA:	3,29	3.2 - 4.8 g/liter
Viscosity (at 25°C)	266,3	> 170 cP
Mouse embryo test (1-cell):	100 %	≥ 80%

Product released by Eileen Vanderplancke on 13/09/2017 (DD/MM/YYYY)

Eileen Vanderplancke

Source materials from which this product was derived was found negative when tested for antibodies against HIV-1/2 and HCV and non-reactive for HbsAg, HBV RNA, HCV RNA, HIV-1 RNA and syphilis. No known test method can offer assurance that products derived from human blood will not transmit infectious agents. Therefore, handle all specimens as if capable of transmitting HIV or hepatitis.