

## Certificate of Analysis and Release

Product: SpermFreeze

Batchnumber:		FP17SF05
Production date:	DD/MM/YYYY	8/03/2017
Expiry date:	DD/MM/YYYY	30/09/2018
Total number of kits:		87 kits

	Results	Specification
pH:	7,57	7.20-7.60
Bacterial screen:	PASSED	PASSED
Fungal screen:	PASSED	PASSED
Sperm survival Test (4 hours exposure):	98%%	≥ 80%
Endotoxin:	< 0,05 EU/ml	< 0,25 EU/mL
Concentration HSA:	3,689	3,2 - 4,8 g/liter

Product released by Justine Van Renterghem on 23/03/2017 (DD/MM/YYYY)

