

Certificate of Analysis and Release

Product: SpermFreeze

Batchnumber:		FP17SF10
Production date:	DD/MM/YYYY	12/06/2017
Expiry date:	DD/MM/YYYY	31/12/2018
Total number of kits:		85 kits

	Results	Specification
pH:	7,53	7.20-7.60
Bacterial screen:	PASSED	PASSED
Fungal screen:	PASSED	PASSED
Sperm survival Test (4 hours exposure):	94 %	≥ 80%
Endotoxin:	< 0,05 EU/ml	< 0,25 EU/mL
Concentration HSA:	4,1	3,2 - 4,8 g/liter

Product released by Eileen Vanderplancke on 27/06/2017 (DD/MM/YYYY)

Eileen Vanderplancke