

Certificate of Analysis and Release

Product: SpermFreeze

Batchnumber:		FP17SF13
Production date:	DD/MM/YYYY	14/07/2017
Expiry date:	DD/MM/YYYY	31/01/2019
Total number of kits:		68 kits

	Results	Specification
pH:	7,48	7.20-7.60
Bacterial screen:	PASSED	PASSED
Fungal screen:	PASSED	PASSED
Sperm survival Test (4 hours exposure):	93 %	≥ 80%
Endotoxin:	< 0,05 EU/mL	< 0,25 EU/mL
Concentration HSA:	3,653	3,2 - 4,8 g/liter

Product released by Eileen Vanderplancke on 31/07/2017 (DD/MM/YYYY)

Eileen Vanderplancke