

Certificate of Analysis and Release

Product: Sil-Select Plus

| | | |
|-------------------|------------|------------|
| Batchnumber : | | FP17SIP09L |
| Production date : | DD/MM/YYYY | 6/03/2017 |
| Expiry date : | DD/MM/YYYY | 30/09/2018 |
| Total volume : | | 34130 ml |

| Sil-Select Plus: Upper layer | <i>Result</i> | <i>Specification</i> |
|-------------------------------------|---------------|----------------------|
| pH : | 7,41 | 7.20 - 7.60 |
| Osmolality : | 323 | 310-340mOsm/kg |
| Bacterial screen : | NEGATIVE | NEGATIVE |
| Fungal screen | NEGATIVE | NEGATIVE |
| Endotoxin : | | < 0,5EU/mL |
| MEA 1: | | ≥ 80% |
| Sil-Select Plus: Lower layer | | |
| pH : | 7,41 | 7.20 - 7.60 |
| Osmolality : | 326 | 320-350mOsm/kg |
| Bacterial screen : | NEGATIVE | NEGATIVE |
| Fungal screen | NEGATIVE | NEGATIVE |
| Endotoxin : | < 0,1 EU/MI | < 0,5EU/mL |
| Sperm survival test: | 95% | ≥ 80% |
| Density (g/ml): | 1,1075 | 1.1050 - 1.1150 g/mL |
| Viscosity in cP (25°C): | 1,45 | < 1.65 cP |

Product released by Justine Van Renterghem on 21/03/2017 (DD/MM/YYYY)

