

Certificate of Analysis and Release

Product: Sil-Select Plus - Lower Layer

Batchnumber :		FP17SIP11L
Production date :	DD/MM/YYYY	21/03/2017
Expiry date :	DD/MM/YYYY	30/09/2018
Total volume :		27500 ml

<i>Sil-select Plus: Lower layer</i>	<i>Result</i>	<i>Specification</i>
pH :	7,48	7.20 - 7.60
Osmolality :	338	320-350mOsm/kg
Bacterial screen :	NEGATIVE	NEGATIVE
Fungal screen	NEGATIVE	NEGATIVE
Endotoxin :	< 0,1 EU/mL	< 0,5EU/mL
Density (g/ml):	1,1086	1.1050 - 1.1150 g/mL
Sperm survival test:	96%	≥ 80%
Viscosity in cP (25°C):	1,44	< 1.65 cP

Product released by Justine Van Renterghem on 05/04/2017 (DD/MM/YYYY)

