

Certificate of Analysis and Release

Product: Sil-Select Plus - Upper Layer

Batchnumber :		FP17SIP11U
Production date :	DD/MM/YYYY	21/03/2017
Expiry date :	DD/MM/YYYY	30/09/2018
Total volume :		23000 ml

<i>Sil-select Plus: Upper layer</i>	<i>Result</i>	<i>Specification</i>
pH :	7,37	7.20 - 7.60
Osmolality :	318	310-340mOsm/kg
Bacterial screen :	NEGATIVE	NEGATIVE
Fungal screen	NEGATIVE	NEGATIVE
Endotoxin :	<0,1 EU/ml	< 0,5EU/mL
Sperm survival test	97%	≥ 80%

Product released by Justine Van Renterghem on 05/04/2017 (DD/MM/YYYY)

