

Certificate of Analysis and Release

Product: Sil-Select Plus - Upper Layer

Batchnumber :		FP17SIP13U
Production date :	DD/MM/YYYY	3/04/2017
Expiry date :	DD/MM/YYYY	31/10/2018
Total volume :		23450 ml

Sil-select Plus: Upper layer

	<i>Result</i>	<i>Specification</i>
pH :	7,39	7.20 - 7.60
Osmolality :	324	310-340mOsm/kg
Bacterial screen :	NEGATIVE	NEGATIVE
Fungal screen	NEGATIVE	NEGATIVE
Endotoxin :	< 0,1 EU/ml	< 0,5EU/mL
Sperm survival test	97%	≥ 80%

Product released by Justine Van Renterghem on 18/04/2017 (DD/MM/YYYY)

