

## Certificate of Analysis and Release

Product: Sil-Select Plus - Lower Layer

Batchnumber :		FP17SIP16L
Production date :	DD/MM/YYYY	18/04/2017
Expiry date :	DD/MM/YYYY	31/10/2018
Total volume :		10000 ml

**Sil-select Plus: Lower layer**

	<i>Result</i>	<i>Specification</i>
pH :	7,44	7.20 - 7.60
Osmolality :	344	320-350mOsm/kg
Bacterial screen :	NEGATIVE	NEGATIVE
Fungal screen	NEGATIVE	NEGATIVE
Endotoxin :	< 0,1 EU/mL	< 0,5EU/mL
Density (g/ml):	1,1107	1.1050 - 1.1150 g/mL
Sperm survival test (4 hours exposure)	100%	≥ 80%
Viscosity in cP (25°C):	1,49	< 1.65 cP

Product released by Justine Van Renterghem on 08/05/2017 (DD/MM/YYYY)

