

Certificate of Analysis and Release

Product: Sil-Select Plus - Upper Layer

Batchnumber :		FP17SIP16U
Production date :	DD/MM/YYYY	18/04/2017
Expiry date :	DD/MM/YYYY	31/10/2018
Total volume :		10000 ml

Sil-select Plus: Upper layer

	<i>Result</i>	<i>Specification</i>
pH :	7,46	7.20 - 7.60
Osmolality :	322	310-340mOsm/kg
Bacterial screen :	NEGATIVE	NEGATIVE
Fungal screen	NEGATIVE	NEGATIVE
Endotoxin :	< 0,1 EU/ml	< 0,5EU/mL
Sperm survival test (4 hours exposure)	100 %	≥ 80%

Product released by Justine Van Renterghem on 08/05/2017 (DD/MM/YYYY)

A handwritten signature in blue ink, appearing to read "Justine Van Renterghem". The signature is written in a cursive style and is underlined with a horizontal line.