

Certificate of Analysis and Release

Product: Sil-Select Plus - Lower Layer

Batchnumber :		FP17SIP17L
Production date :	DD/MM/YYYY	4/05/2017
Expiry date :	DD/MM/YYYY	30/11/2018
Total volume :		23850 ml

Sil-select Plus: Lower layer

	<i>Result</i>	<i>Specification</i>
pH :	7,51	7.20 - 7.60
Osmolality :	334	320-350mOsm/kg
Bacterial screen :	NEGATIVE	NEGATIVE
Fungal screen :	NEGATIVE	NEGATIVE
Endotoxin :	< 0.1 EU/mL	< 0,5EU/mL
Density (g/ml):	1,1096	1.1050 - 1.1150 g/mL
Sperm survival test (4 hours exposure)	100 %	≥ 80%
Viscosity in cP (25°C):	1,49	< 1.65 cP

Product released by Eileen Vanderplancke on 19/05/2017 (DD/MM/YYYY)

Eileen Vanderplancke