

Certificate of Analysis and Release

Product: Sil-Select Plus - Upper Layer

Batchnumber :		FP17SIP17U
Production date :	DD/MM/YYYY	4/05/2017
Expiry date :	DD/MM/YYYY	30/11/2018
Total volume :		13200 ml

Sil-select Plus: Upper layer

	<i>Result</i>	<i>Specification</i>
pH :	7,49	7.20 - 7.60
Osmolality :	320	310-340mOsm/kg
Bacterial screen :	NEGATIVE	NEGATIVE
Fungal screen :	NEGATIVE	NEGATIVE
Endotoxin :	< 0,1 EU/ml	< 0,5EU/mL
Sperm survival test	97%	≥ 80%

Product released by Eileen Vanderplancke on 19/05/2017 (DD/MM/YYYY)

Eileen Vanderplancke