

## Certificate of Analysis and Release

Product: Sil-Select Plus - Lower Layer

Batchnumber :		FP17SIP19L
Production date :	DD/MM/YYYY	1/06/2017
Expiry date :	DD/MM/YYYY	31/12/2018
Total volume :		50200 ml

**Sil-select Plus: Lower layer**

	<i>Result</i>	<i>Specification</i>
pH :	7,42	7.20 - 7.60
Osmolality :	336	320-350mOsm/kg
Bacterial screen :	NEGATIVE	NEGATIVE
Fungal screen	NEGATIVE	NEGATIVE
Endotoxin :	0,352 EU/ml	< 0,5EU/mL
Density (g/ml):	1,1085	1.1050 - 1.1150 g/mL
Sperm survival test (4 hours exposure)	97 %	≥ 80%
Viscosity in cP (25°C):	1,56	< 1.65 cP

Product released by Justine Van Renterghem on 21/06/2017 (DD/MM/YYYY)

