

Certificate of Analysis and Release

Product: Sil-Select Plus - Upper Layer

Batchnumber :		FP17SIP19U
Production date :	DD/MM/YYYY	1/06/2017
Expiry date :	DD/MM/YYYY	31/12/2018
Total volume :		30100 ml

Sil-select Plus: Upper layer

	<i>Result</i>	<i>Specification</i>
pH :	7,43	7.20 - 7.60
Osmolality :	318	310-340mOsm/kg
Bacterial screen :	NEGATIVE	NEGATIVE
Fungal screen	NEGATIVE	NEGATIVE
Endotoxin :	< 0,05 EU/ml	< 0,5EU/mL
Sperm survival test (4 hours exposure)	97 %	≥ 80%

Product released by Justine Van Renterghem on 21/06/2017 (DD/MM/YYYY)

A handwritten signature in blue ink, appearing to read "Justine Van Renterghem", with a horizontal line drawn underneath it.