

Certificate of Analysis and Release

Product: Sil-Select Plus - Lower Layer

Batchnumber :		FP17SIP24L
Production date :	DD/MM/YYYY	3/07/2017
Expiry date :	DD/MM/YYYY	31/01/2019
Total volume :		24050 ml

Sil-select Plus: Lower layer	<i>Result</i>	<i>Specification</i>
pH :	7,40	7.20 - 7.60
Osmolality :	329	320-350mOsm/kg
Bacterial screen :	NEGATIVE	NEGATIVE
Fungal screen :	NEGATIVE	NEGATIVE
Endotoxin :	0,337 EU/mL	< 0,5EU/mL
Density (g/ml):	1,1071	1.1050 - 1.1150 g/mL
Sperm survival test (4 hours exposure)	95 %	≥ 80%
Viscosity in cP (25°C):	1,44	< 1.65 cP

Product released by Eileen Vanderplancke on 18/07/2017 (DD/MM/YYYY)

Eileen Vanderplancke