

Certificate of Analysis and Release

Product: Sil-Select Plus - Upper Layer

Batchnumber :		FP17SIP26U
Production date :	DD/MM/YYYY	13/07/2017
Expiry date :	DD/MM/YYYY	31/01/2019
Total volume :		35250 ml

<i>Sil-select Plus: Upper layer</i>	<i>Result</i>	<i>Specification</i>
pH :	7,48	7.20 - 7.60
Osmolality :	320	310-340mOsm/kg
Bacterial screen :	NEGATIVE	NEGATIVE
Fungal screen :	NEGATIVE	NEGATIVE
Endotoxin :	< 0,2 EU/ml	< 0,5EU/mL
Sperm survival test (4 hours exposure)	91	≥ 80%

Product released by Eileen Vanderplancke on 31/07/0207 (DD/MM/YYYY)

Eileen Vanderplancke