

## Certificate of Analysis and Release

Product: Sil-Select Plus - Lower Layer

Batchnumber :		FP17SIP32L
Production date :	DD/MM/YYYY	6/09/2017
Expiry date :	DD/MM/YYYY	31/03/2019
Total volume :		20000 ml

<b>Sil-select Plus: Lower layer</b>	<i>Result</i>	<i>Specification</i>
pH :	7,44	7.20 - 7.60
Osmolality :	321	320-350mOsm/kg
Bacterial screen :	NEGATIVE	NEGATIVE
Fungal screen :	NEGATIVE	NEGATIVE
Endotoxin :	< 0.2 EU/MI	< 0,5EU/mL
Density (g/ml):	1,1051	1.1050 - 1.1150 g/mL
Sperm survival test (4 hours exposure)	100%	≥ 80%
Viscosity in cP (25 °C):	1,46	< 1.65 cP

Product released by Lenie Bocklandt on 22/09/2017 (DD/MM/YYYY)



Bocklandt denie