

## Certificate of Analysis and Release

### Product: Sil-Select Plus - Upper Layer

Batchnumber :		FP17SIP33U
Production date :	DD/MM/YYYY	8/09/2017
Expiry date :	DD/MM/YYYY	31/03/2019
Total volume :		10000 ml

<b><i>Sil-select Plus: Upper layer</i></b>	<i>Result</i>	<i>Specification</i>
pH :	7,45	7.20 - 7.60
Osmolality :	324	310-340mOsm/kg
Bacterial screen :	NEGATIVE	NEGATIVE
Fungal screen :	NEGATIVE	NEGATIVE
Endotoxin :	< 0,1 EU/ml	< 0,5EU/mL
Sperm survival test (4 hours exposure)	98%	≥ 80%

Product released by Lenie Bocklandt on 25/09/2017 (DD/MM/YYYY)



Bocklandt denie