

Certificate of Analysis and Release

Product: Sil-Select Plus - Lower Layer

Batchnumber :		FP17SIP44L
Production date :	DD/MM/YYYY	30/11/2017
Expiry date :	DD/MM/YYYY	31/05/2019
Total volume :		22440 ml

Sil-select Plus: Lower layer

	<i>Result</i>	<i>Specification</i>
pH :	7,45	7.20 - 7.60
Osmolality :	323	320-350mOsm/kg
Bacterial screen :	NEGATIVE	NEGATIVE
Fungal screen	NEGATIVE	NEGATIVE
Endotoxin :	< 0.1 EU/mL	< 0,5EU/mL
Density (g/ml):	1,1082	1.1050 - 1.1150 g/mL
Sperm survival test (4 hours exposure)	96%	≥ 80%
Viscosity in cP (25°C):	1,48	< 1.65 cP

Product released by Justine Van Renterghem on 15/12/2017 (DD/MM/YYYY)

