

## Certificate of Analysis and Release

### Product: Sil-Select Plus - Upper Layer

Batchnumber :		FP17SIP44U
Production date :	DD/MM/YYYY	30/11/2017
Expiry date :	DD/MM/YYYY	31/05/2019
Total volume :		11600 ml

#### **Sil-select Plus: Upper layer**

	<i>Result</i>	<i>Specification</i>
pH :	7,41	7.20 - 7.60
Osmolality :	310	310-340mOsm/kg
Bacterial screen :	NEGATIVE	NEGATIVE
Fungal screen	NEGATIVE	NEGATIVE
Endotoxin :	< 0,1 EU/ml	< 0,5EU/mL
Sperm survival test (4 hours exposure)	94%	≥ 80%

Product released by Justine Van Renterghem on 15/12/2017 (DD/MM/YYYY)

