

Certificate of Analysis and Release

Product: Sil-Select Plus - Upper Layer

Batchnumber :		FP17SIP45U
Production date :	DD/MM/YYYY	12/12/2017
Expiry date :	DD/MM/YYYY	30/06/2019
Total volume :		21600 ml

<i>Sil-select Plus: Upper layer</i>	<i>Result</i>	<i>Specification</i>
pH :	7,52	7.20 - 7.60
Osmolality :	324	310-340mOsm/kg
Bacterial screen :	NEGATIVE	NEGATIVE
Fungal screen	NEGATIVE	NEGATIVE
Endotoxin :	0,217 EU/ml	< 0,5EU/mL
Sperm survival test (4 hours exposure)	100%	≥ 80%

Product released by Justine Van Renterghem on 27/12/2017 (DD/MM/YYYY)

A handwritten signature in blue ink, appearing to be "Justine Van Renterghem", written over a horizontal line.