

Certificate of Analysis and Release

Product: Sil-Select Stock

Batchnumber :		FP17SIS08
Production date :	DD/MM/YYYY	2/05/2017
Expiry date :	DD/MM/YYYY	30/11/2018
Total volume :		12400 ml
	<i>Result</i>	<i>Specification</i>
pH :	7,47	7.20 - 7.60
Osmolality :	317	300 - 330 mOsm/kg
Sterility test bacteria:	NEGATIVE	NEGATIVE
Sterility test fungi:	NEGATIVE	NEGATIVE
Endotoxin:	< 0,1 EU/ml	< 0,5 EU/mL
Sperm survival test (4 hours exposure)	99 %	≥ 80%
Density (g/mL):	1,1220	1,1150-1,1250 g/mL
Viscosity in cP (25°C):	1,56	< 1,75 cP

Product released by Eileen Vanderplancke on 17/05/2017 (DD/MM/YYYY)

Eileen Vanderplancke

