

Certificate of Analysis and Release

Product: Sil-Select Stock

Batchnumber :		FP17SIS16
Production date :	DD/MM/YYYY	18/09/2017
Expiry date :	DD/MM/YYYY	31/03/2019
Total volume :		10200 ml

	<i>Result</i>	<i>Specification</i>
pH :	7,54	7.20 - 7.60
Osmolality :	312	300 - 330 mOsm/kg
Sterility test bacteria:	NEGATIVE	NEGATIVE
Sterility test fungi:	NEGATIVE	NEGATIVE
Endotoxin:	0,108 EU/ml	< 0,5 EU/mL
Sperm survival test (4 hours exposure)	100 %	≥ 80%
Density (g/mL):	1,1201	1,1150-1,1250 g/mL
Viscosity in cP (25°C):	1,60	< 1,75 cP

Product released by Eileen Vanderplancke on 03/10/2017 (DD/MM/YYYY)

Eileen Vanderplancke