

Certificate of Analysis and Release

Product: VitriThaw ES media

Batchnumber :		FP17VTE06
Production date :	DD/MM/YYYY	24/04/2017
Expiry date :	DD/MM/YYYY	30/04/2018
Total batchsize:		200 Kits

VitriThaw ES medium 5	Result	Specification
pH :	7,30	7.2 - 7.4
Osmolality :	282	270-295mOsm/kg
Sterility test :	PASSED	PASSED
Endotoxine :	<0,100 EU/mL	< 0,25EU/mL
Concentration HSA:	19,86	16,0 - 24,0 g/liter

VitriThaw ES medium 4		
pH :	7,28	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin :	<0,1	< 0,25EU/mL
Osmolality :	412	405-435mOsm/kg
Concentration HSA:	18,951	15.5 - 23.3 g/liter

VitriThaw ES medium 3		
pH :	7,34	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin :	0,102	< 0,25EU/mL
Osmolality :	559	535-565mOsm/kg
Concentration HSA :	18,321	15.2 - 22.8 g/liter

VitriThaw ES medium 2		
pH :	7,28	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin :	<0,100 EU/mL	< 0,25EU/mL
Osmolality :	848	805-850mOsm/kg
Concentration HSA :	17,731	14.5 - 21.7 g/liter

VitriThaw ES medium 1		
pH :	7,21	7.2-7.4
Sterility test :	PASSED	PASSED
Endotoxin :	<0,100 EU/mL	< 0,25EU/mL
Concentration HSA :	16,389	13.2 - 19.8 g/liter
Mouse Embryo Assay (blastocysts after 96h)	100%	≥ 80%

Product released by Justine Van Renterghem on 09/05/2017 (DD/MM/YYYY)

