

Certificate of Analysis and Release

Product: VitriThaw ES media

Batchnumber :		FP17VTE11
Production date :	DD/MM/YYYY	9/08/2017
Expiry date :	DD/MM/YYYY	31/08/2018
Total batchsize:		186 Kits

VitriThaw ES medium 5	Result	Specification
pH :	7,25	7.2 - 7.4
Osmolality :	283	270-295mOsm/kg
Sterility test :	PASSED	PASSED
Endotoxine :	0,109 Eu/mL	< 0,25EU/mL
Concentration HSA:	19,407	16,0 - 24,0 g/liter

VitriThaw ES medium 4		
pH :	7,31	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin :	< 0,100 EU/mL	< 0,25EU/mL
Osmolality :	418	405-435mOsm/kg
Concentration HSA:	18,076	15.5 - 23.3 g/liter

VitriThaw ES medium 3		
pH :	7,31	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin :	< 0,100 EU/mL	< 0,25EU/mL
Osmolality :	558	535-565mOsm/kg
Concentration HSA :	17,558	15.2 - 22.8 g/liter

VitriThaw ES medium 2		
pH :	7,37	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin :	< 0,1EU/mL	< 0,25EU/mL
Osmolality :	834	805-850mOsm/kg
Concentration HSA :	16,258	14.5 - 21.7 g/liter

VitriThaw ES medium 1		
pH :	7,25	7.2-7.4
Sterility test :	PASSED	PASSED
Endotoxin :	< 0,100 EU/mL	< 0,25EU/mL
Concentration HSA :	15,705	13.2 - 19.8 g/liter

Mouse Embryo Assay (blastocysts after 96h)	95 %	≥ 80%
--	------	-------

Product released by Eileen Vanderplancke on 24/08/2017 (DD/MM/YYYY)

Eileen Vanderplancke

Caution: All blood products should be treated as potentially infectious. Source material from which this product was derived was found negative when tested for antibodies to HIV, HBc, HCV, and non-reactive for HbsAg, HCV RNS and HIV-1 RNA and syphilis. No known test methods can offer assurance that products derived from human blood will not transmit infectious agents.