

Certificate of Analysis and Release

Product: VitriThaw ES media

Batchnumber :		FP17VTE15
Production date :	DD/MM/YYYY	16/10/2017
Expiry date :	DD/MM/YYYY	31/10/2018
Total batchsize:		199 Kits

VitriThaw ES medium 5

	Result	Specification
pH :	7,30	7.2 - 7.4
Osmolality :	281	270-295mOsm/kg
Sterility test :	PASSED	PASSED
Endotoxine :	0,116 EU/mL	< 0,25EU/mL
Concentration HSA:	18,644	16,0 - 24,0 g/liter

VitriThaw ES medium 4

pH :	7,28	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin :	0,208 EU/mL	< 0,25EU/mL
Osmolality :	418	405-435mOsm/kg
Concentration HSA:	19,903	15.5 - 23.3 g/liter

VitriThaw ES medium 3

pH :	7,36	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin :	0,105 EU/mL	< 0,25EU/mL
Osmolality :	555	535-565mOsm/kg
Concentration HSA :	18,982	15.2 - 22.8 g/liter

VitriThaw ES medium 2

pH :	7,30	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin :	0,152 EU/mL	< 0,25EU/mL
Osmolality :	833	805-850mOsm/kg
Concentration HSA :	17,592	14.5 - 21.7 g/liter

VitriThaw ES medium 1

pH :	7,32	7.2-7.4
Sterility test :	PASSED	PASSED
Endotoxin :	< 0,100 EU/mL	< 0,25EU/mL
Concentration HSA :	16,649	13.2 - 19.8 g/liter

Mouse Embryo Assay (blastocysts after 96h)	100 %	≥ 80%
--	-------	-------

Product released by Eileen Vanderplancke on 31/10/2017 (DD/MM/YYYY)

Eileen Vanderplancke

Caution: All blood products should be treated as potentially infectious. Source material from which this product was derived was found negative when tested for antibodies to HIV, HBc, HCV, and non-reactive for HbsAg, HCV RNS and HIV-1 RNA and syphilis. No known test methods can offer assurance that products derived from human blood will not transmit infectious agents.