

Certificate of Analysis and Release

Product: VitriThaw ES media

Batchnumber :		FP17VTE16
Production date :	DD/MM/YYYY	29/11/2017
Expiry date :	DD/MM/YYYY	30/11/2018
Total batchsize:		200 Kits

VitriThaw ES medium 5	Result	Specification
pH :	7,28	7.2 - 7.4
Osmolality :	283	270-295mOsm/kg
Sterility test :	PASSED	PASSED
Endotoxine :	<0,1 EU/ml	< 0,25EU/mL
Concentration HSA:	19,59	16,0 - 24,0 g/liter

VitriThaw ES medium 4		
pH :	7,32	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin :	<0,1 EU/ml	< 0,25EU/mL
Osmolality :	413	405-435mOsm/kg
Concentration HSA:	19,04	15.5 - 23.3 g/liter

VitriThaw ES medium 3		
pH :	7,34	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin :	<0,1 EU/ml	< 0,25EU/mL
Osmolality :	553	535-565mOsm/kg
Concentration HSA :	18,801	15.2 - 22.8 g/liter

VitriThaw ES medium 2		
pH :	7,30	7.2 - 7.4
Sterility test :	PASSED	PASSED
Endotoxin :	<0,1 EU/ml	< 0,25EU/mL
Osmolality :	841	805-850mOsm/kg
Concentration HSA :	16,159	14.5 - 21.7 g/liter

VitriThaw ES medium 1		
pH :	7,29	7.2-7.4
Sterility test :	PASSED	PASSED
Endotoxin :	<0,1 EU/ml	< 0,25EU/mL
Concentration HSA :	16,322	13.2 - 19.8 g/liter

Mouse Embryo Assay (blastocysts after 96h)	95,24%	≥ 80%
--	--------	-------

Product released by Justine Van Renterghem on 14/12/2017 (DD/MM/YYYY)

