

Certificate of Analysis and Release

Product: SpermMar IgA

Batchnumber :		FP18A01
Production date :	DD/MM/YY	8/01/2018
Expiry date :	DD/MM/YY	31/01/2019
Number of kits in batch:		245 kits

Chemical analysis

pH:		8,58
Osmolality:		338
Concentration:		PASSED

Functional analysis

Positive control:		PASSED
Negative control:		PASSED

Product released by Lenie Bocklandt on 17/01/2018 (DD/MM/YYYY)

A handwritten signature in blue ink, appearing to be "Lenie Bocklandt". Below the signature, the name "Bocklandt denie" is written in a cursive, handwritten style.