

Certificate of Analysis and Release

Product: Embryo Freeze with propanediol


Batchnumber :		FP18EF02
Production date :	DD/MM/YYYY	23/03/2018
Expiry date :	DD/MM/YYYY	30/09/2019
Total batchsize:		19 Kits

Freezing medium	Results	Specifications
pH :	7,22	7.20-7.40
Bacterial screen :	PASSED	PASSED
Fungal screen :	PASSED	PASSED
Endotoxine :	<0,1 EU/ml	< 0.25EU/mL
Concentration HSA:	15,84	11,7 - 17,6 g/liter
Thawing medium 1		
pH :	7,23	7.20-7.40
Bacterial screen :	PASSED	PASSED
Fungal screen :	PASSED	PASSED
Endotoxin :	<0,1 EU/ml	< 0.25EU/mL
Concentration HSA:	15,152	11,4 - 17,1 g/liter
Thawing medium 2		
pH :	7,23	7.20-7.40
Bacterial screen :	PASSED	PASSED
Fungal screen :	PASSED	PASSED
Endotoxin :	<0,1 EU/ml	< 0.25EU/mL
Concentration HSA:	15,202	11,4 - 17,1 g/liter
Thawing medium 3		
pH :	7,23	7.20-7.40
Osmolality:	543	530-560mOsm/kg
Bacterial screen :	PASSED	PASSED
Fungal screen :	PASSED	PASSED
Endotoxin :	<0,1 EU/ml	< 0.25EU/mL
Concentration HSA:	14,782	11,4 - 17,1 g/liter

All media

MEA (1-cell to expanded blastocyst within 96 hours):	95%	≥ 80%
--	-----	-------

Product released by Lenie Bocklandt on 09/04/2018 (DD/MM/YYYY)



Bocklandt lenie

*The plasma from Talecris (Grifols) which is the source for our human albumin is tested for anti-HIV 1, anti-HIV 2, anti-HCV and HBsAG.
The albumin responds to all the conditions for the FDA and the EMEA regarding safety.*

*The plasma from Talecris (Grifols) which is the source for our human albumin is tested for anti-HIV 1, anti-HIV 2, anti-HCV and HBsAG.
The albumin responds to all the conditions for the FDA and the EMEA regarding safety.*