

Certificate of Analysis and Release

Product: EpiScreen Plus

Batchnumber:		FP18EP01
Production date:	DD/MM/YYYY	9/01/2018
Expiry date:	DD/MM/YYYY	31/01/2020
Batch size:		94 kits

	<i>Result</i>	<i>Specification</i>
--	---------------	----------------------

Reagent 1

Osmolality (mOsm/kg):	232	210-260 mOsm/kg
pH:	6,78	6,70 - 6,90

Reagent 3

Osmolality (mOsm/kg):	1027	930-1140 mOsm/kg
pH:	6,63	6,55 - 6,75

Reagent 4

pH:	12,02	> 12
-----	-------	------

Reagent 6

pH:	12,08	> 12
-----	-------	------

Functionality test:	PASSED	PASSED
---------------------	--------	--------

Product released by Justine Van Renterghem on 19/01/2018 (DD/MM/YYYY)

