

Certificate of Analysis and Release

Product: EpiScreen Plus

Batchnumber:		FP18EP02
Production date:	DD/MM/YYYY	2/03/2018
Expiry date:	DD/MM/YYYY	31/03/2020
Batch size:		100 kits

	<i>Result</i>	<i>Specification</i>
Reagent 1		
Osmolality (mOsm/kg):	230	210-260 mOsm/kg
pH:	6,74	6,70 - 6,90
Reagent 3		
Osmolality (mOsm/kg):	1006	930-1140 mOsm/kg
pH:	6,67	6,55 - 6,75
Reagent 4		
pH:	12,05	> 12
Reagent 6		
pH:	12,01	> 12
Functionality test:	PASSED	PASSED

Product released by Lenie Bocklandt on 14/03/2018 (DD/MM/YYYY)



Bocklandt lenie