

Certificate of Analysis and Release

Product: EpiScreen Plus

Batchnumber:		FP18EP03
Production date:	DD/MM/YYYY	13/06/2018
Expiry date:	DD/MM/YYYY	30/06/2020
Batch size:		100 kits

	<i>Result</i>	<i>Specification</i>
Reagent 1		
Osmolality (mOsm/kg):	232	210-260 mOsm/kg
pH:	6,80	6,70 - 6,90
Reagent 3		
Osmolality (mOsm/kg):	1020	930-1140 mOsm/kg
pH:	6,65	6,55 - 6,75
Reagent 4		
pH:	12,12	> 12
Reagent 6		
pH:	12,16	> 12
Functionality test:	PASSED	PASSED

Product released by Lenie Bocklandt on 22/06/2018 (DD/MM/YYYY)



Bocklandt lenie